
Midmark M11 - v1.0.0
Total Cycles: 2445

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2022 07:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.2 F 0.0
2:00 94.0 F 0.0
4:00 88.8 F 0.0
6:00 118.0 F 1.9
8:00 157.1 F 4.8
10:00 177.8 F 8.2
12:00 194.8 F 11.8
14:00 215.7 F 3.7
16:00 223.4 F 7.4
18:00 234.8 F 11.3
20:00 231.6 F 7.0
22:00 243.1 F 12.8
24:00 239.9 F 10.6
26:00 252.0 F 16.7
28:00 263.4 F 23.0
30:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 271.6 F 28.8
2:00 272.1 F 28.9
3:00 272.4 F 29.1
4:00 272.4 F 29.4

Min 271.3 F 28.8
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 213.9 F 0.0
10:00 196.4 F 0.0
15:00 193.2 F 0.0
20:00 202.5 F 0.0
25:00 203.6 F 0.0
30:00 199.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:57
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:06
CYCLE COMPLETE

10/18/2022 08:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____