
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 005144
12:51 21/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:26
STERILIZING 3:31
136.4 C 220kPa 3:31
136.3 C 222kPa 4:01
136.4 C 225kPa 4:31
136.4 C 222kPa 5:01
135.9 C 220kPa 5:31
136.4 C 224kPa 6:01
136.4 C 225kPa 6:31
136.4 C 223kPa 7:01
136.4 C 223kPa 7:01
VENTING 7:01
AIR DRYING 7:52
CYCLE COMPLETE 7:59

Digital Signature #
DD80E5C5B11E3D6F
