

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2332

Name: RITTER2  
Sterilizer ID: V2313228  
05/03/2023 12:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.1 F 0.0  
2:00 188.4 F 4.2  
4:00 202.6 F 9.7  
6:00 219.8 F 7.4  
8:00 232.1 F 11.5  
10:00 233.6 F 9.3  
12:00 237.2 F 9.5  
14:00 248.2 F 15.1  
16:00 260.7 F 22.1  
18:00 271.7 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.2 F 29.7  
2:00 271.8 F 29.9  
3:00 271.9 F 29.5  
4:00 272.0 F 30.1  
  
Min 271.4 F 29.2  
Max 272.4 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.4 F 0.0  
5:00 211.8 F 0.0  
10:00 222.3 F 0.0  
15:00 227.4 F 0.0  
20:00 231.1 F 0.0  
25:00 229.9 F 0.0  
30:00 227.9 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 18:32  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 00:57:50  
CYCLE COMPLETE

05/03/2023 01:36 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_