
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002696
10:02 17/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:38
STERILIZING 3:42
136.5 C 224kPa 3:42
136.3 C 224kPa 4:12
135.9 C 217kPa 4:42
136.4 C 220kPa 5:12
136.4 C 223kPa 5:42
136.2 C 223kPa 6:12
136.4 C 223kPa 6:42
136.4 C 222kPa 7:12
136.4 C 222kPa 7:12
VENTING 7:12
AIR DRYING 8:06
CYCLE COMPLETE 8:11

Digital Signature #
4C5D2CE8FED31A17
