

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4136

Name: RITTER 1  
Sterilizer ID: V2178198  
12/19/2023 02:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 93.0 F 0.0  
2:00 93.6 F 0.6  
4:00 106.2 F 2.5  
6:00 133.0 F 5.2  
8:00 159.5 F 8.5  
10:00 190.8 F 12.3  
12:00 210.0 F 4.8  
14:00 222.0 F 8.4  
16:00 232.7 F 12.4  
18:00 230.6 F 8.0  
20:00 242.0 F 13.0  
22:00 239.9 F 10.6  
24:00 251.6 F 16.3  
26:00 261.8 F 22.2  
28:00 271.5 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 271.8 F 29.0  
2:00 271.9 F 29.0  
3:00 272.1 F 29.0  
4:00 272.4 F 29.2

Min 271.7 F 29.0  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.9 F 0.0  
5:00 223.2 F 0.0  
10:00 235.9 F 0.0  
15:00 217.8 F 0.0  
20:00 221.1 F 0.0  
25:00 227.5 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 29:24  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:08:01  
CYCLE COMPLETE

12/19/2023 03:36 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_