

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4433

Name: RITTER 1  
Sterilizer ID: V2178198  
04/08/2024 09:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss      Degrees      PSI  
0:00      67.6 F      0.0

C103: HEATUP MODE  
STOP PRESSED

\*\* ITEMS NOT STERILE \*\*

04/08/2024 09:31 AM