

-----  
Midmark M11 - v1.0.0  
Total Cycles: 641

Name: MIDMARK  
Sterilizer ID: V2178198  
01/28/2021 12:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.6 F 0.0  
2:00 183.8 F 3.9  
4:00 201.7 F 9.3  
6:00 224.5 F 5.3  
8:00 235.0 F 11.8  
10:00 236.6 F 9.3  
12:00 236.7 F 8.5  
14:00 250.8 F 15.4  
16:00 263.9 F 23.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.0  
1:00 272.0 F 28.1  
2:00 272.3 F 28.5  
3:00 272.0 F 28.7  
4:00 271.4 F 27.9  
  
Min 271.0 F 27.9  
Max 272.5 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.2 F 0.0  
5:00 239.2 F 0.0  
10:00 233.9 F 0.0  
15:00 228.9 F 0.0  
20:00 227.1 F 0.0  
25:00 227.6 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 1:57  
HEATING: 17:47  
STERILIZING: 4:00  
VENTING: 2:56  
DRYING: 30:00  
TOTAL CYCLE: 00:56:40  
CYCLE COMPLETE

01/28/2021 01:04 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_