
Midmark M11 - v1.0.5
Total Cycles: 1461

Name: RITTER2
Sterilizer ID: V2313228
09/07/2022 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.4 F 0.0
2:00 132.8 F 0.0
4:00 150.0 F 0.0
6:00 172.5 F 2.8
8:00 192.6 F 6.1
10:00 209.0 F 10.3
12:00 222.1 F 7.0
14:00 230.2 F 9.7
16:00 236.3 F 10.0
18:00 241.4 F 12.1
20:00 239.9 F 10.7
22:00 252.4 F 16.9
24:00 263.1 F 23.2
26:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.6 F 29.7
2:00 272.4 F 29.3
3:00 271.6 F 28.7
4:00 271.6 F 29.2

Min 271.3 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.8 F 0.0
5:00 210.7 F 0.0
10:00 225.7 F 0.0
15:00 231.3 F 0.0
20:00 228.6 F 0.0
25:00 228.3 F 0.0
30:00 224.8 F 0.0

Drying Complete

FILLING: 0:53
HEATING: 26:20
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:04:14
CYCLE COMPLETE

09/07/2022 11:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____