
Midmark M11 - v1.0.0
Total Cycles: 4795

Name: RITTER 1
Sterilizer ID: V2178198
08/08/2024 11:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.0 F 0.0
2:00 153.4 F 0.6
4:00 175.3 F 3.5
6:00 194.0 F 7.2
8:00 225.2 F 11.2
10:00 218.9 F 4.6
12:00 230.5 F 8.6
14:00 240.1 F 12.9
16:00 237.8 F 9.6
18:00 239.3 F 9.8
20:00 246.8 F 13.3
22:00 257.7 F 19.2
24:00 267.6 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.1
1:00 271.5 F 28.3
2:00 272.3 F 28.9
3:00 272.4 F 28.8
4:00 272.1 F 28.5

Min 271.2 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 232.6 F 0.0
10:00 217.1 F 0.0
15:00 193.7 F 0.0
20:00 204.7 F 0.0
25:00 210.1 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 25:31
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:04:11
CYCLE COMPLETE

08/08/2024 12:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____