
Midmark M11 - v1.0.0
Total Cycles: 4645

Name: RITTER 1
Sterilizer ID: V2178198
06/14/2024 03:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.0 F 0.0
2:00 112.2 F 1.1
4:00 135.2 F 3.3
6:00 159.0 F 6.1
8:00 177.4 F 9.1
10:00 193.5 F 12.6
12:00 213.3 F 5.0
14:00 219.1 F 8.6
16:00 229.7 F 12.4
18:00 232.1 F 7.9
20:00 241.2 F 12.6
22:00 239.9 F 10.3
24:00 251.6 F 16.0
26:00 261.7 F 21.9
28:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 271.7 F 28.6
2:00 272.2 F 28.8
3:00 272.1 F 29.0
4:00 272.1 F 29.2

Min 271.3 F 28.6
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 233.7 F 0.0
10:00 187.9 F 0.0
15:00 186.2 F 0.0
20:00 183.6 F 0.0
25:00 182.2 F 0.0
30:00 180.4 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 29:17
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:08:03
CYCLE COMPLETE

06/14/2024 04:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____