
Midmark M11 - v1.0.5
Total Cycles: 2985

Name: RITTER2
Sterilizer ID: V2313228
10/10/2023 08:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 72.1 F 0.0
4:00 91.9 F 0.3
6:00 126.3 F 2.0
8:00 158.8 F 5.1
10:00 186.8 F 8.9
12:00 217.8 F 12.5
14:00 218.1 F 5.0
16:00 218.1 F 8.7
18:00 230.6 F 12.7
20:00 231.0 F 8.4
22:00 241.6 F 13.4
24:00 240.5 F 11.4
26:00 252.2 F 17.4
28:00 263.1 F 23.7
30:00 272.7 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.4
1:00 272.2 F 29.8
2:00 271.2 F 29.7
3:00 272.1 F 29.7
4:00 271.8 F 29.8

Min 271.0 F 29.2
Max 273.1 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 205.1 F 0.0
10:00 201.7 F 0.0
15:00 199.4 F 0.0
20:00 201.9 F 0.0
25:00 209.7 F 0.0
30:00 214.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:19
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:21
CYCLE COMPLETE

10/10/2023 09:22 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____