

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000817  
11:37 10/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:03  
PRESSURIZING 3:15  
STERILIZING 4:23  
136.0 C 227kPa 4:23  
136.4 C 231kPa 4:53  
136.2 C 231kPa 5:23  
136.4 C 233kPa 5:53  
136.4 C 233kPa 6:23  
136.4 C 233kPa 6:53  
136.4 C 233kPa 7:23  
136.2 C 230kPa 7:53  
136.2 C 230kPa 7:53  
VENTING 7:53  
AIR DRYING 8:46  
CYCLE COMPLETE 23:46  
-----

Digital Signature #  
82F0B4B76E87572F  
-----