

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2532

Name: RITTER 1  
Sterilizer ID: V2178198  
11/08/2022 04:28 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.3 F 0.0  
2:00 172.0 F 3.7  
4:00 180.5 F 8.2  
6:00 203.1 F 13.3  
8:00 221.9 F 7.5  
10:00 236.3 F 12.7  
12:00 235.6 F 9.6  
14:00 240.1 F 8.7  
16:00 248.2 F 14.2  
18:00 259.9 F 20.8  
20:00 271.0 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.4 F 29.0  
2:00 271.6 F 28.3  
3:00 271.6 F 28.7  
4:00 272.2 F 28.9  
  
Min 271.2 F 28.1  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.3 F 0.0  
5:00 234.4 F 0.0  
10:00 228.0 F 0.0  
15:00 229.2 F 0.0  
20:00 228.9 F 0.0  
25:00 230.5 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 20:45  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:59  
CYCLE COMPLETE

11/08/2022 05:26 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_