

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3806

Name: RITTER2  
Sterilizer ID: V2313228  
07/22/2024 03:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 95.2 F 0.0  
2:00 95.9 F 0.0  
4:00 104.0 F 0.1  
6:00 145.7 F 0.3  
8:00 163.6 F 3.2  
10:00 191.7 F 6.5  
12:00 208.4 F 10.1  
14:00 221.7 F 10.2  
16:00 226.0 F 9.1  
18:00 236.6 F 13.5  
20:00 237.4 F 10.7  
22:00 234.9 F 8.3  
24:00 248.0 F 14.8  
26:00 258.7 F 20.7  
28:00 268.5 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.9  
1:00 271.5 F 30.2  
2:00 271.9 F 30.2  
3:00 272.3 F 30.2  
4:00 271.9 F 30.1  
  
Min 271.3 F 29.7  
Max 272.7 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.9 F 0.0  
5:00 216.6 F 0.0  
10:00 219.2 F 0.0  
15:00 216.9 F 0.0  
20:00 220.0 F 0.0  
25:00 227.9 F 0.0  
30:00 228.3 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 29:50  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:08:29  
CYCLE COMPLETE

07/22/2024 04:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_