
Midmark M11 - v1.0.5
Total Cycles: 2569

Name: RITTER2
Sterilizer ID: V2313228
06/26/2023 02:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.3 F 0.0
2:00 189.1 F 3.4
4:00 198.7 F 8.2
6:00 216.1 F 11.1
8:00 228.8 F 9.5
10:00 234.6 F 9.4
12:00 243.0 F 13.2
14:00 242.8 F 12.3
16:00 255.7 F 19.0
18:00 266.9 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 271.9 F 29.6
2:00 272.5 F 29.5
3:00 271.5 F 29.0
4:00 272.2 F 29.8

Min 271.3 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 218.6 F 0.0
10:00 225.0 F 0.0
15:00 224.3 F 0.0
20:00 225.2 F 0.0
25:00 225.4 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 19:21
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

06/26/2023 03:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____