
Midmark M11 - v1.0.0
Total Cycles: 1267

Name: RITTER 1
Sterilizer ID: V2178198
10/14/2021 11:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.0 F 0.0
2:00 184.4 F 3.8
4:00 195.5 F 8.2
6:00 217.7 F 11.3
8:00 231.6 F 8.0
10:00 241.0 F 13.3
12:00 240.4 F 10.8
14:00 237.2 F 9.1
16:00 252.1 F 16.2
18:00 264.1 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.7 F 28.9
2:00 271.6 F 28.1
3:00 272.1 F 28.8
4:00 272.4 F 28.7

Min 271.2 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 218.2 F 0.0
10:00 218.2 F 0.0
15:00 227.7 F 0.0
20:00 215.0 F 0.0
25:00 217.8 F 0.0
30:00 183.9 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:59
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

10/14/2021 12:46 PM
ACCEPTED: X
REJECTED:
Operator: U02