
Midmark M11 - v1.0.0
Total Cycles: 2784

Name: RITTER 1
Sterilizer ID: V2178198
01/17/2023 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.1 F 0.0
2:00 173.7 F 2.9
4:00 180.3 F 6.7
6:00 207.3 F 12.1
8:00 220.6 F 6.6
10:00 233.7 F 11.8
12:00 233.8 F 8.7
14:00 240.1 F 10.2
16:00 246.1 F 13.7
18:00 258.8 F 20.5
20:00 270.3 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.4 F 29.2
2:00 271.8 F 29.1
3:00 272.2 F 29.0
4:00 271.8 F 29.1

Min 271.5 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 230.5 F 0.0
10:00 231.3 F 0.0
15:00 232.2 F 0.0
20:00 228.5 F 0.0
25:00 227.6 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 20:50
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

01/17/2023 12:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____