
Midmark M11 - v1.0.5
Total Cycles: 3180

Name: RITTER2
Sterilizer ID: V2313228
12/04/2023 12:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.3 F 0.0
2:00 135.2 F 1.1
4:00 157.7 F 4.3
6:00 180.7 F 8.2
8:00 200.6 F 12.8
10:00 220.7 F 7.3
12:00 230.8 F 12.0
14:00 232.2 F 9.1
16:00 238.5 F 10.3
18:00 245.0 F 13.6
20:00 257.2 F 20.0
22:00 267.9 F 26.8
24:00 271.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.8
1:00 272.5 F 30.0
2:00 273.0 F 30.3
3:00 271.6 F 30.0
4:00 272.2 F 30.0

Min 271.5 F 29.7
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.3 F 0.0
5:00 211.0 F 0.0
10:00 194.3 F 0.0
15:00 193.3 F 0.0
20:00 192.3 F 0.0
25:00 191.2 F 0.0
30:00 190.4 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:04
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:02:37
CYCLE COMPLETE

12/04/2023 01:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____