
Midmark M11 - v1.0.0
Total Cycles: 2997

Name: RITTER 1
Sterilizer ID: V2178198
03/13/2023 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.8 F 0.0
2:00 183.7 F 3.7
4:00 203.2 F 8.0
6:00 223.6 F 11.7
8:00 229.5 F 8.7
10:00 238.7 F 10.9
12:00 242.1 F 12.0
14:00 240.6 F 10.8
16:00 254.6 F 17.8
18:00 266.5 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.6 F 28.9
2:00 271.2 F 28.2
3:00 272.8 F 29.1
4:00 271.7 F 28.4

Min 271.1 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 225.8 F 0.0
10:00 241.3 F 0.0
15:00 237.6 F 0.0
20:00 234.2 F 0.0
25:00 223.8 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

03/13/2023 02:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____