
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001406
14:58 25/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:30
PRESSURIZING 4:30
STERILIZING 5:36
136.0 C 221kPa 5:36
136.2 C 222kPa 6:07
136.3 C 223kPa 6:37
136.4 C 225kPa 7:07
136.4 C 225kPa 7:37
136.3 C 224kPa 8:07
136.2 C 224kPa 8:37
136.4 C 226kPa 9:07
136.4 C 226kPa 9:07
VENTING 9:07
AIR DRYING 10:11
CYCLE COMPLETE 10:32

Digital Signature #
506E839F3F03CEAB
