
Midmark M11 - v1.0.5
Total Cycles: 1567

Name: RITTER2
Sterilizer ID: V2313228
10/05/2022 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.6 F 0.0
2:00 139.0 F 0.0
4:00 158.3 F 0.0
6:00 170.9 F 0.2
8:00 196.9 F 1.8
10:00 217.8 F 5.9
12:00 227.0 F 10.6
14:00 232.0 F 7.6
16:00 242.9 F 12.5
18:00 240.4 F 11.0
20:00 239.1 F 10.3
22:00 253.1 F 17.2
24:00 264.8 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.3
2:00 271.7 F 29.0
3:00 272.6 F 29.3
4:00 271.5 F 28.7

Min 271.5 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 215.9 F 0.0
10:00 237.8 F 0.0
15:00 232.4 F 0.0
20:00 224.6 F 0.0
25:00 222.4 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 25:50
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:03:54
CYCLE COMPLETE

10/05/2022 02:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____