

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001589  
17:07 13/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:08  
PRESSURIZING 2:13  
STERILIZING 3:14  
136.0 C 220kPa 3:14  
136.0 C 222kPa 3:45  
136.3 C 224kPa 4:15  
136.4 C 226kPa 4:45  
136.1 C 223kPa 5:15  
136.2 C 224kPa 5:45  
136.2 C 225kPa 6:15  
136.4 C 226kPa 6:45  
136.4 C 226kPa 6:45  
VENTING 6:45  
AIR DRYING 7:34  
CYCLE COMPLETE 7:40  
-----

Digital Signature #  
CBE0CB7E0FEB2DE0  
-----