
Midmark M11 - v1.0.0
Total Cycles: 3281

Name: RITTER 1
Sterilizer ID: V2178198
05/25/2023 07:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.9 F 0.0
2:00 80.9 F 0.0
4:00 95.4 F 0.0
6:00 122.4 F 0.0
8:00 156.1 F 0.7
10:00 184.1 F 4.1
12:00 214.3 F 7.9
14:00 225.0 F 11.6
16:00 225.4 F 4.7
18:00 234.7 F 9.2
20:00 243.4 F 11.9
22:00 240.4 F 10.5
24:00 236.1 F 8.0
26:00 249.6 F 14.6
28:00 260.8 F 20.9
30:00 271.2 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 27.9
1:00 271.5 F 28.4
2:00 272.3 F 28.3
3:00 271.1 F 27.8
4:00 272.1 F 28.7

Min 271.0 F 27.7
Max 272.4 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 218.0 F 0.0
10:00 226.3 F 0.0
15:00 232.9 F 0.0
20:00 226.6 F 0.0
25:00 230.9 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:35
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:08:34
CYCLE COMPLETE

05/25/2023 08:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____