

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001138  
9:49 23/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:18  
PRESSURIZING 3:20  
STERILIZING 4:30  
136.0 C 221kPa 4:30  
136.4 C 225kPa 5:00  
136.3 C 225kPa 5:30  
136.3 C 225kPa 6:00  
136.2 C 222kPa 6:30  
136.4 C 226kPa 7:00  
136.2 C 224kPa 7:30  
136.2 C 222kPa 8:00  
136.2 C 222kPa 8:00  
VENTING 8:00  
AIR DRYING 8:52  
CYCLE COMPLETE 17:06  
-----

Digital Signature #  
0B60EF0A90DC9764  
-----