
Midmark M11 - v1.0.5
Total Cycles: 3887

Name: RITTER2
Sterilizer ID: V2313228
08/23/2024 08:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 70.5 F 0.0
4:00 94.4 F 0.0
6:00 112.1 F 0.4
8:00 148.3 F 2.3
10:00 160.6 F 5.3
12:00 184.9 F 8.4
14:00 195.1 F 11.9
16:00 211.2 F 4.6
18:00 221.1 F 8.1
20:00 230.7 F 12.0
22:00 231.3 F 8.1
24:00 242.0 F 13.1
26:00 241.2 F 11.3
28:00 252.1 F 16.9
30:00 262.2 F 22.8
32:00 271.8 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.5 F 29.8
2:00 271.8 F 29.0
3:00 271.7 F 29.5
4:00 272.4 F 29.6

Min 271.2 F 28.8
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 199.3 F 0.0
10:00 201.6 F 0.0
15:00 202.4 F 0.0
20:00 213.5 F 0.0
25:00 217.4 F 0.0
30:00 211.9 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 32:38
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:11:01
CYCLE COMPLETE

08/23/2024 09:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____