

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4271

Name: RITTER 1  
Sterilizer ID: V2178198  
02/05/2024 03:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.8 F 0.0  
2:00 145.8 F 1.9  
4:00 148.1 F 5.0  
6:00 182.3 F 8.3  
8:00 193.3 F 12.6  
10:00 217.5 F 4.9  
12:00 223.0 F 8.9  
14:00 233.7 F 13.2  
16:00 236.6 F 9.6  
18:00 238.8 F 9.8  
20:00 246.5 F 13.4  
22:00 257.7 F 19.4  
24:00 267.8 F 25.8  
26:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.7 F 29.2  
2:00 271.8 F 29.0  
3:00 272.6 F 29.1  
4:00 271.7 F 29.0

Min 271.6 F 28.7  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.8 F 0.0  
5:00 229.8 F 0.0  
10:00 227.3 F 0.0  
15:00 228.3 F 0.0  
20:00 227.4 F 0.0  
25:00 228.9 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 26:13  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:05:12  
CYCLE COMPLETE

02/05/2024 04:34 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_