
Midmark M11 - v1.0.0
Total Cycles: 3318

Name: RITTER 1
Sterilizer ID: V2178198
06/01/2023 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.7 F 0.0
2:00 186.1 F 0.0
4:00 193.7 F 2.1
6:00 218.5 F 7.4
8:00 226.5 F 12.9
10:00 231.5 F 8.6
12:00 242.2 F 11.3
14:00 243.3 F 12.2
16:00 241.9 F 11.1
18:00 255.4 F 18.0
20:00 267.1 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.8 F 28.9
2:00 272.0 F 28.3
3:00 271.4 F 28.2
4:00 272.9 F 29.0

Min 271.3 F 27.9
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 224.2 F 0.0
10:00 222.8 F 0.0
15:00 224.7 F 0.0
20:00 221.8 F 0.0
25:00 219.9 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:45
CYCLE COMPLETE

06/01/2023 09:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____