
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001189
11:48 27/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:46
PRESSURIZING 4:47
STERILIZING 6:13
136.0 C 220kPa 6:13
136.4 C 224kPa 6:43
136.4 C 223kPa 7:13
136.4 C 222kPa 7:43
136.4 C 224kPa 8:13
136.1 C 222kPa 8:43
136.4 C 223kPa 9:13
136.4 C 224kPa 9:43
136.4 C 224kPa 9:43
VENTING 9:44
AIR DRYING 10:43
CYCLE COMPLETE 25:43

Digital Signature #
F3C673A9C29A72EF
