
Midmark M11 - v1.0.0
Total Cycles: 1296

Name: RITTER 1
Sterilizer ID: V2178198
10/25/2021 10:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.7 F 0.0
2:00 180.8 F 1.9
4:00 189.0 F 5.7
6:00 209.3 F 10.4
8:00 219.0 F 5.6
10:00 231.2 F 9.9
12:00 237.6 F 9.2
14:00 242.8 F 12.4
16:00 240.9 F 10.8
18:00 253.4 F 17.1
20:00 264.3 F 23.5
22:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.5 F 28.8
2:00 271.9 F 28.6
3:00 272.4 F 28.8
4:00 271.7 F 28.5

Min 271.6 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.5 F 0.0
5:00 214.5 F 0.0
10:00 223.0 F 0.0
15:00 215.9 F 0.0
20:00 211.5 F 0.0
25:00 218.3 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 22:01
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:01:11
CYCLE COMPLETE

10/25/2021 11:08 AM
ACCEPTED: X
REJECTED:
Operator: U01