
Midmark M11 - v1.0.0
Total Cycles: 1413

Name: RITTER 1
Sterilizer ID: V2178198
12/01/2021 09:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.7 F 0.0
2:00 157.8 F 0.9
4:00 159.9 F 4.1
6:00 187.5 F 8.5
8:00 207.1 F 13.4
10:00 222.3 F 7.3
12:00 235.6 F 12.4
14:00 235.4 F 9.3
16:00 239.8 F 9.4
18:00 247.0 F 13.8
20:00 259.3 F 20.5
22:00 270.5 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 29.1
2:00 271.6 F 28.3
3:00 272.2 F 29.0
4:00 272.3 F 28.9

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 214.7 F 0.0
10:00 225.2 F 0.0
15:00 221.9 F 0.0
20:00 221.9 F 0.0
25:00 222.7 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 22:51
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

12/01/2021 10:41 AM
ACCEPTED: X
REJECTED:
Operator: U02