
Midmark M11 - v1.0.0
Total Cycles: 4017

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2023 10:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.8 F 0.0
2:00 182.4 F 2.3
4:00 182.1 F 6.0
6:00 194.8 F 10.3
8:00 222.2 F 4.8
10:00 221.6 F 8.1
12:00 234.8 F 12.9
14:00 235.7 F 9.9
16:00 237.6 F 9.0
18:00 247.9 F 14.4
20:00 259.5 F 20.8
22:00 270.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.3
2:00 273.0 F 29.2
3:00 272.3 F 28.8
4:00 271.7 F 28.5

Min 271.2 F 28.4
Max 273.2 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 233.7 F 0.0
10:00 223.9 F 0.0
15:00 218.2 F 0.0
20:00 219.4 F 0.0
25:00 218.1 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 23:02
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:01
CYCLE COMPLETE

11/14/2023 11:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____