

-----  
Midmark M11 - v1.0.0  
Total Cycles: 711

Name: MIDMARK  
Sterilizer ID: V2178198  
03/01/2021 09:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 124.5 F 0.0  
2:00 129.7 F 0.4  
4:00 135.4 F 2.5  
6:00 160.0 F 6.1  
8:00 193.0 F 10.7  
10:00 221.8 F 4.1  
12:00 225.9 F 9.6  
14:00 235.9 F 8.7  
16:00 242.1 F 12.6  
18:00 240.7 F 10.7  
20:00 254.0 F 17.4  
22:00 265.7 F 24.5  
24:00 272.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 271.9 F 29.2  
2:00 272.4 F 29.2  
3:00 272.3 F 29.1  
4:00 271.8 F 29.0

Min 271.2 F 28.7  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.0 F 0.0  
5:00 231.4 F 0.0  
10:00 226.6 F 0.0  
15:00 233.9 F 0.0  
20:00 228.4 F 0.0  
25:00 222.1 F 0.0  
30:00 205.4 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 24:09  
STERILIZING: 4:00  
VENTING: 2:56  
DRYING: 30:00  
TOTAL CYCLE: 01:01:57  
CYCLE COMPLETE

03/01/2021 10:26 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01