

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1391

Name: RITTER 1  
Sterilizer ID: V2178198  
11/24/2021 09:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.4 F 0.0  
2:00 92.8 F 0.0  
4:00 90.0 F 0.0  
6:00 115.3 F 0.1  
8:00 152.1 F 0.2  
10:00 179.8 F 0.6  
12:00 201.2 F 3.8  
14:00 217.2 F 7.2  
16:00 226.7 F 11.0  
18:00 227.3 F 6.4  
20:00 236.6 F 10.0  
22:00 237.7 F 9.1  
24:00 245.8 F 13.1  
26:00 243.8 F 11.8  
28:00 256.6 F 18.4  
30:00 267.7 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.9  
1:00 272.6 F 28.5  
2:00 271.6 F 27.9  
3:00 272.7 F 28.6  
4:00 271.8 F 27.9

Min 271.5 F 27.8  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.1 F 0.0  
5:00 233.5 F 0.0  
10:00 216.3 F 0.0  
15:00 209.0 F 0.0  
20:00 198.4 F 0.0  
25:00 201.5 F 0.0  
30:00 205.3 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:21  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:09:21  
CYCLE COMPLETE

11/24/2021 10:41 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_