

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1547

Name: RITTER 1  
Sterilizer ID: V2178198  
01/20/2022 01:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.1 F 0.0  
2:00 180.7 F 3.7  
4:00 190.2 F 8.6  
6:00 213.0 F 8.6  
8:00 224.5 F 8.7  
10:00 235.5 F 10.8  
12:00 239.8 F 11.3  
14:00 238.4 F 9.5  
16:00 252.1 F 16.3  
18:00 263.6 F 23.1  
20:00 271.6 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.4 F 28.7  
2:00 271.8 F 28.6  
3:00 272.3 F 28.7  
4:00 271.5 F 28.4  
  
Min 271.5 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.0 F 0.0  
5:00 223.2 F 0.0  
10:00 233.3 F 0.0  
15:00 228.4 F 0.0  
20:00 229.6 F 0.0  
25:00 229.9 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 1:36  
HEATING: 20:04  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:54  
CYCLE COMPLETE

01/20/2022 02:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_