

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 007736  
11:01 13/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:47  
PRESSURIZING 3:47  
STERILIZING 4:44  
136.4 C 222kPa 4:44  
136.3 C 224kPa 5:14  
136.4 C 225kPa 5:44  
136.3 C 223kPa 6:14  
136.3 C 223kPa 6:44  
136.3 C 221kPa 7:14  
136.3 C 224kPa 7:44  
136.4 C 225kPa 8:14  
136.4 C 225kPa 8:14  
VENTING 8:14  
AIR DRYING 9:14  
CYCLE COMPLETE 15:10

-----  
Digital Signature #  
66C667DB07E39AB5  
-----