
Midmark M11 - v1.0.0
Total Cycles: 4949

Name: RITTER 1
Sterilizer ID: V2178198
10/01/2024 02:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.3 F 0.0
2:00 158.0 F 0.2
4:00 176.7 F 1.0
6:00 199.8 F 5.0
8:00 214.5 F 9.5
10:00 230.1 F 9.6
12:00 230.9 F 9.6
14:00 235.6 F 9.1
16:00 242.7 F 12.7
18:00 241.2 F 11.4
20:00 254.5 F 18.0
22:00 266.0 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.6 F 29.2
2:00 271.9 F 28.6
3:00 272.1 F 29.0
4:00 272.4 F 29.1

Min 271.4 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 225.6 F 0.0
10:00 224.3 F 0.0
15:00 227.9 F 0.0
20:00 227.5 F 0.0
25:00 229.3 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 23:41
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:09
CYCLE COMPLETE

10/01/2024 03:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____