
Midmark M11 - v1.0.0
Total Cycles: 1058

Name: RITTER 1
Sterilizer ID: V2178198
07/27/2021 08:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.9 F 0.0
2:00 185.4 F 2.8
4:00 198.8 F 7.5
6:00 217.5 F 12.8
8:00 228.3 F 8.3
10:00 241.3 F 11.9
12:00 241.2 F 11.6
14:00 239.8 F 10.1
16:00 253.4 F 16.9
18:00 264.8 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.1 F 28.6
2:00 272.4 F 29.0
3:00 271.4 F 28.2
4:00 272.6 F 28.9

Min 271.2 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 237.4 F 0.0
10:00 229.1 F 0.0
15:00 225.9 F 0.0
20:00 226.5 F 0.0
25:00 228.2 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

07/27/2021 09:59 AM
ACCEPTED: X
REJECTED:
Operator: U01