
Midmark M11 - v1.0.0
Total Cycles: 4318

Name: RITTER 1
Sterilizer ID: V2178198
02/21/2024 04:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.1 F 0.0
2:00 167.9 F 2.8
4:00 176.6 F 6.4
6:00 209.6 F 10.7
8:00 221.9 F 4.5
10:00 230.1 F 9.4
12:00 237.8 F 9.8
14:00 238.4 F 11.3
16:00 237.6 F 9.3
18:00 250.3 F 15.7
20:00 261.8 F 22.2
22:00 272.1 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 271.7 F 29.1
2:00 272.5 F 29.2
3:00 271.4 F 28.6
4:00 271.3 F 28.9

Min 271.0 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 206.2 F 0.0
10:00 213.5 F 0.0
15:00 212.3 F 0.0
20:00 206.9 F 0.0
25:00 203.7 F 0.0
30:00 202.1 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 22:25
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:01:39
CYCLE COMPLETE

02/21/2024 05:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____