
Midmark M11 - v1.0.0
Total Cycles: 2323

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2022 09:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.1 F 0.0
2:00 150.9 F 2.0
4:00 170.3 F 5.5
6:00 184.9 F 9.7
8:00 209.7 F 5.8
10:00 218.0 F 6.8
12:00 230.8 F 11.5
14:00 231.7 F 7.7
16:00 243.4 F 13.4
18:00 242.4 F 11.6
20:00 254.9 F 18.0
22:00 265.8 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.7
1:00 272.5 F 29.4
2:00 271.5 F 28.7
3:00 272.5 F 29.3
4:00 271.4 F 28.7

Min 271.4 F 28.6
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 233.5 F 0.0
10:00 219.6 F 0.0
15:00 227.5 F 0.0
20:00 232.2 F 0.0
25:00 237.8 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 23:37
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:38
CYCLE COMPLETE

09/14/2022 10:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____