
Midmark M11 - v1.0.0
Total Cycles: 2522

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2022 04:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.8 F 0.0
2:00 153.4 F 2.9
4:00 174.2 F 6.6
6:00 196.0 F 11.1
8:00 215.1 F 4.7
10:00 227.4 F 9.6
12:00 233.4 F 9.3
14:00 240.2 F 11.5
16:00 237.5 F 9.4
18:00 251.3 F 16.0
20:00 262.8 F 22.6
22:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.0
2:00 271.6 F 28.6
3:00 272.4 F 28.9
4:00 271.4 F 28.5

Min 271.4 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 225.9 F 0.0
10:00 240.9 F 0.0
15:00 234.5 F 0.0
20:00 229.9 F 0.0
25:00 221.5 F 0.0
30:00 215.7 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:30
CYCLE COMPLETE

11/07/2022 05:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____