
Midmark M11 - v1.0.5
Total Cycles: 2814

Name: RITTER2
Sterilizer ID: V2313228
08/29/2023 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.2 F 0.0
2:00 179.9 F 1.3
4:00 203.9 F 6.2
6:00 221.8 F 11.4
8:00 228.8 F 7.9
10:00 241.4 F 12.8
12:00 242.1 F 12.3
14:00 242.1 F 11.7
16:00 255.6 F 18.7
18:00 267.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.0 F 29.0
2:00 272.6 F 29.5
3:00 271.8 F 29.3
4:00 272.0 F 29.0

Min 271.5 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 216.8 F 0.0
10:00 216.4 F 0.0
15:00 216.6 F 0.0
20:00 212.4 F 0.0
25:00 210.3 F 0.0
30:00 207.9 F 0.0

Drying Complete

FILLING: 2:56
HEATING: 19:20
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:35
CYCLE COMPLETE

08/29/2023 05:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____