
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002069
10:03 20/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:32
PRESSURIZING 1:38
STERILIZING 2:36
136.4 C 222kPa 2:36
136.3 C 225kPa 3:06
136.4 C 225kPa 3:36
136.3 C 226kPa 4:06
136.4 C 224kPa 4:36
136.4 C 226kPa 5:06
136.3 C 226kPa 5:36
136.2 C 225kPa 6:06
136.2 C 225kPa 6:06
VENTING 6:06
AIR DRYING 6:54
CYCLE COMPLETE 15:25

Digital Signature #
556DC82944C3D03E
