

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1144

Name: RITTER2  
Sterilizer ID: V2313228  
05/31/2022 10:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.4 F 0.0  
2:00 197.2 F 3.2  
4:00 194.6 F 8.5  
6:00 212.5 F 9.4  
8:00 225.6 F 9.4  
10:00 233.9 F 9.8  
12:00 240.9 F 12.5  
14:00 240.5 F 11.4  
16:00 253.3 F 17.8  
18:00 264.5 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.4  
1:00 272.5 F 29.7  
2:00 271.5 F 29.7  
3:00 272.4 F 29.7  
4:00 271.4 F 29.5  
  
Min 271.3 F 29.2  
Max 272.6 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.8 F 0.0  
5:00 208.8 F 0.0  
10:00 224.9 F 0.0  
15:00 223.2 F 0.0  
20:00 226.4 F 0.0  
25:00 223.9 F 0.0  
30:00 214.1 F 0.0  
Drying Complete

FILLING: 2:05  
HEATING: 19:56  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:11  
CYCLE COMPLETE

05/31/2022 11:07 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02