

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3340

Name: RITTER 1  
Sterilizer ID: V2178198  
06/06/2023 08:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.9 F 0.0  
2:00 191.5 F 0.0  
4:00 201.9 F 3.8  
6:00 224.2 F 9.2  
8:00 227.6 F 8.5  
10:00 236.6 F 10.6  
12:00 236.6 F 7.6  
14:00 242.7 F 11.4  
16:00 245.7 F 13.1  
18:00 258.0 F 19.7  
20:00 269.1 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.6 F 28.9  
2:00 271.5 F 28.4  
3:00 272.5 F 28.9  
4:00 271.6 F 28.3  
  
Min 271.5 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 220.2 F 0.0  
10:00 215.4 F 0.0  
15:00 220.5 F 0.0  
20:00 231.8 F 0.0  
25:00 230.6 F 0.0  
30:00 227.7 F 0.0  
Drying Complete

FILLING: 1:45  
HEATING: 21:06  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:00:06  
CYCLE COMPLETE

06/06/2023 09:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_