

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006590  
9:24 10/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:03  
STERILIZING 3:03  
136.5 C 220kPa 3:03  
136.3 C 223kPa 3:33  
136.2 C 222kPa 4:03  
136.4 C 224kPa 4:33  
136.4 C 222kPa 5:03  
136.4 C 224kPa 5:33  
136.4 C 222kPa 6:03  
136.4 C 223kPa 6:33  
136.4 C 223kPa 6:33  
VENTING 6:33  
AIR DRYING 7:23  
CYCLE COMPLETE 37:23

-----  
Digital Signature #  
C41199C7D8DD144F  
-----