

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2407

Name: RITTER 1  
Sterilizer ID: V2178198  
10/06/2022 07:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.8 F 0.0  
2:00 77.2 F 0.0  
4:00 85.2 F 0.2  
6:00 114.1 F 1.4  
8:00 145.7 F 3.1  
10:00 173.1 F 5.8  
12:00 191.6 F 9.0  
14:00 206.4 F 12.5  
16:00 222.9 F 5.6  
18:00 234.6 F 9.6  
20:00 237.1 F 10.5  
22:00 239.3 F 10.7  
24:00 234.7 F 8.0  
26:00 249.3 F 14.7  
28:00 260.5 F 20.9  
30:00 270.9 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.7 F 28.7  
2:00 272.3 F 29.0  
3:00 271.8 F 28.4  
4:00 272.4 F 29.1  
  
Min 271.3 F 28.3  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.7 F 0.0  
5:00 224.4 F 0.0  
10:00 221.6 F 0.0  
15:00 249.7 F 0.0  
20:00 224.5 F 0.0  
25:00 231.3 F 0.0  
30:00 222.5 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 31:30  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:09:32  
CYCLE COMPLETE

10/06/2022 08:12 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_