
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001284
10:01 04/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:19
PRESSURIZING 3:21
STERILIZING 4:25
136.4 C 223kPa 4:25
136.4 C 224kPa 4:55
136.4 C 225kPa 5:25
136.2 C 222kPa 5:55
136.4 C 225kPa 6:25
136.3 C 222kPa 6:55
136.1 C 222kPa 7:25
136.4 C 225kPa 7:55
136.4 C 225kPa 7:55
VENTING 7:55
AIR DRYING 8:46
CYCLE COMPLETE 23:47

Digital Signature #
54F66CE4F7F9B06D
