
Midmark M11 - v1.0.0
Total Cycles: 2428

Name: RITTER 1
Sterilizer ID: V2178198
10/13/2022 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.7 F 0.0
2:00 176.7 F 3.9
4:00 199.1 F 8.5
6:00 225.0 F 13.3
8:00 228.0 F 7.9
10:00 240.6 F 13.3
12:00 240.1 F 10.8
14:00 236.7 F 9.0
16:00 251.1 F 15.9
18:00 263.1 F 22.8
20:00 272.1 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.6 F 29.0
2:00 271.2 F 28.5
3:00 272.6 F 29.1
4:00 271.3 F 28.4

Min 271.2 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.9 F 0.0
5:00 231.9 F 0.0
10:00 224.7 F 0.0
15:00 229.4 F 0.0
20:00 228.7 F 0.0
25:00 229.2 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 20:10
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

10/13/2022 12:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____