
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007624
7:25 03/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:51
STERILIZING 3:53
136.4 C 221kPa 3:53
136.3 C 218kPa 4:23
136.1 C 221kPa 4:53
136.3 C 221kPa 5:23
136.4 C 223kPa 5:53
136.3 C 220kPa 6:23
136.0 C 221kPa 6:53
136.2 C 219kPa 7:23
136.2 C 219kPa 7:23
VENTING 7:23
AIR DRYING 8:13
CYCLE COMPLETE 8:21

Digital Signature #
006216EA76AFE9AB
