
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000705
7:11 20/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:05
STERILIZING 3:11
136.4 C 220kPa 3:11
136.4 C 232kPa 3:41
136.3 C 229kPa 4:11
136.1 C 230kPa 4:41
136.4 C 233kPa 5:11
136.4 C 229kPa 5:41
136.2 C 230kPa 6:11
136.3 C 230kPa 6:41
136.3 C 230kPa 6:41
VENTING 6:41
AIR DRYING 7:28
CYCLE COMPLETE 7:53

Digital Signature #
E5AE160EBC7897A3
